



CERTIFICATE OF ANALYSIS

PRODUCT NAME: GLYKOPREP® DIGESTION MODULE (24-ct)
PRODUCT CODE: GS24-RX
LOT NUMBER: 929 028
STORAGE: Store at 2 - 8°C.
PACK SIZE: 2 x 12 cartridges
EXPIRATION: February 2019 (extended from prior exp. date based on re-assay); the components of this module are usable at least through this month.
RE-ASSAY DATE: March 2018

COMPONENTS

Component Name	Lot No. & Expiry	Contents	Lot No.	Expiry
WS0283 Digestion (RX) Cartridges	W170018 Feb 2020			
WS0256-24 Immobilization Reagent Set	W170054 Mar 2019	WS0226-24 Denaturation Reagent, 1 ea (8 ml)	W160105	Mar 2019
		WS0255-24 Blocking Reagent, 1 ea (1.5 ml)	W160135	Mar 2019
WS0259-24 Digestion Reagent Set	W170063 Feb 2019	WS0229 Finishing Reagent, 1 ea (1 ml)	W160171	Jun 2019
		WS0276 25x Digestion Buffer, 1 ea (0.7 ml)	W160172	Mar 2019
		WS0278-24 4x N-Glycanase®, 1 ea (75 µl)	W170005	Feb 2019

QC RELEASE

The components conform to specifications¹.

Authorized Signature

¹ Glycans were released from (50 µg) of human IgG, labeled and cleaned up using randomly selected GlykoPrep components. The labeled glycans were run on a high pressure liquid chromatography column and the relative abundance of major glycan peaks was comparable to approved historical results for the same protein lot.