



## CERTIFICATE OF ANALYSIS

PRODUCT NAME: GLYKOPREP® CLEANUP MODULE (96-ct)  
PRODUCT CODE: GS96-CU  
LOT NUMBER: 990 036  
STORAGE: Store at 2 - 30°C.  
PACK SIZE: 96 cartridges  
EXPIRATION: April 2021; the components of this module are usable at least through this month.

### COMPONENTS

| Part No: | Description             | Quantity     | Lot     | Expiry   |
|----------|-------------------------|--------------|---------|----------|
| WS0263   | Cleanup (CU) Cartridges | 1 rack of 96 | W180017 | Apr 2021 |

### QC RELEASE

The component conforms to specifications<sup>1</sup>.

---

Authorized Signature

- 1 Glycans were released from (50 µg) of human IgG, labeled and cleaned up using randomly selected GlykoPrep components. The labeled glycans were run on a high pressure liquid chromatography column and the relative abundance of major glycan peaks was comparable to approved historical results for the same protein lot.