



CERTIFICATE OF ANALYSIS

PRODUCT NAME: GLYKOPREP® DIGESTION MODULE (96-ct)
PRODUCT CODE: GS96-RX
LOT NUMBER: 969 038
STORAGE: Store at 2 - 8°C.
PACK SIZE: 96 cartridges
EXPIRATION: October 2019; the components of this module are usable at least through this month.

COMPONENTS

Component Name	Lot No. & Expiry	Contents	Lot No.	Expiry
WS0253 Digestion (RX) Cartridges	W180064 Feb 2021			
WS0256 Immobilization Reagent Set	W180032	WS0226 Denaturation Reagent, 1 ea (30 ml)	W170244	Oct 2019
	Oct 2019	WS0255 Blocking Reagent, 1 ea (6 ml)	W180006	Jan 2020
WS0259 Digestion Reagent Set	W180128	WS0229 Finishing Reagent, 2 ea (1 ml)	W170302	Jun 2020
	Dec 2019	WS0276 25x Digestion Buffer, 1 ea (0.7 ml)	W170303	Dec 2019
		WS0278 4x N-Glycanase®, 1 ea (300 µl)	W180070	Sep 2020

QC RELEASE

The components were assayed following the booklet protocol and conform to specifications¹.

Authorized Signature

¹ Glycans were released from (50 µg) of human IgG, labeled and cleaned up using randomly selected GlykoPrep components. The labeled glycans were run on a high pressure liquid chromatography column and the relative abundance of major glycan peaks was comparable to approved historical results for the same protein lot.