



CERTIFICATE OF ANALYSIS

PRODUCT NAME: GLYKOPREP® DIGESTION MODULE (96-ct)
PRODUCT CODE: GS96-RX
LOT NUMBER: 969 039
STORAGE: Store at 2 - 8°C.
PACK SIZE: 96 cartridges
EXPIRATION: October 2019; the components of this module are usable at least through this month.

COMPONENTS

| Component Name | Lot No. & Expiry | Contents | Lot No. | Expiry |
|---|---------------------|--|---------|----------|
| WS0253 Digestion (RX) Cartridges | W180105 Jul 2021 | | | |
| WS0256 Immobilization Reagent Set | W180032 | WS0226 Denaturation Reagent, 1 ea (30 ml) | W170244 | Oct 2019 |
| | Oct 2019 | WS0255 Blocking Reagent, 1 ea (6 ml) | W180006 | Jan 2020 |
| WS0259 Digestion Reagent Set | W180035 | WS0229 Finishing Reagent, 2 ea (1 ml) | W170121 | Nov 2019 |
| | Oct 2019 | WS0276 25x Digestion Buffer, 1 ea (0.7 ml) | W170303 | Dec 2019 |
| | | WS0278 4x N-Glycanase®, 1 ea (300 µl) | W170180 | Oct 2019 |

QC RELEASE

The components were assayed following the booklet protocol and conform to specifications¹.

Authorized Signature

¹ Glycans were released from (50 µg) of human IgG, labeled and cleaned up using randomly selected GlykoPrep components. The labeled glycans were run on a high pressure liquid chromatography column and the relative abundance of major glycan peaks was comparable to approved historical results for the same protein lot.