

CERTIFICATE OF ANALYSIS

PRODUCT NAME: GLYKOPREP® DIGESTION MODULE (96-ct)
 PRODUCT CODE: GS96-RX
 LOT NUMBER: 969 040
 STORAGE: Store at 2 - 8°C.
 PACK SIZE: 96 cartridges
 EXPIRATION: June 2020; the components of this module are usable at least through this month.

COMPONENTS

Component Name	Lot No. & Expiry	Contents	Lot No.	Expiry
WS0253 Digestion (RX) Cartridges	W180105 Jul 2021			
WS0256 Immobilization Reagent Set	W180241	WS0226 Denaturation Reagent, 1 ea (30 ml)	W180217	Oct 2020
	Oct 2020	WS0255 Blocking Reagent, 1 ea (6 ml)	W180218	Oct 2020
WS0259 Digestion Reagent Set	W180242	WS0229 Finishing Reagent, 2 ea (1 ml)	W170302	Jun 2020
	Jun 2020	WS0276 25x Digestion Buffer, 1 ea (0.7 ml)	W180226	Oct 2020
		WS0278 4x N-Glycanase®, 1 ea (300 µl)	W180151	Sep 2020

QC RELEASE

The components were assayed following the booklet protocol and conform to specifications¹.

Authorized Signature

¹ Glycans were released from (50 µg) of human IgG, labeled and cleaned up using randomly selected GlykoPrep components. The labeled glycans were run on a high pressure liquid chromatography column and the relative abundance of major glycan peaks was comparable to approved historical results for the same protein lot.