



CERTIFICATE OF ANALYSIS

PRODUCT NAME: GLYKOPREP®-*plus* CLEANUP MODULE (96-ct)
PRODUCT CODE: GSP-CU
LOT NUMBER: 962 014
STORAGE: Store at room temperature (2 - 30°C).
PACK SIZE: 96 cartridges
EXPIRATION: November 2020; the components of this module are usable at least through this month.

COMPONENTS

Part No:	Description	Quantity	Lot	Expiry
WS0263	Cleanup (CU) Cartridges	1 rack of 96	W180017	Apr 2021
WS0303-GP	5x 2AB Sample Load Solution (12 ml)	1 each	W170205	Nov 2020

QC RELEASE

The components conform to specifications¹.

Authorized Signature

- ¹ Glycans were released from (50 µg) of human IgG, labeled and cleaned up using randomly selected GlykoPrep components. The labeled glycans were run on a high pressure liquid chromatography column and the relative abundance of major glycan peaks was comparable to approved historical results for the same protein lot.