



## CERTIFICATE OF ANALYSIS

PRODUCT NAME: GLYKOPREP®-*plus* DIGESTION MODULE (96-ct)  
PRODUCT CODE: GSP-RX  
LOT NUMBER: 959 025  
STORAGE: Store at 2 - 8°C.  
PACK SIZE: 96 cartridges  
EXPIRATION: October 2019; the components of this module are usable at least through this month.

### COMPONENTS

Component Name	Lot No. & Expiry	Contents	Lot No.	Expiry
WS0253 Digestion (RX) Cartridges	W180105 Jul 2021			
WS0256-GP Immobilization Reagent Set	W180147 Jan 2020	WS0226-GP Denaturation Reagent 1 ea (35ml)	W180029	Feb 2020
		WS0255-GP Blocking Reagent 1 ea (14 ml)	W180006	Jan 2020
WS0259-GP Digestion Reagent Set	W180146 Oct 2019	WS0229-GP Finishing Reagent 1 ea (10 ml)	W170121	Nov 2019
		WS0276-GP 25x Digestion Buffer 1 ea (1.3 ml)	W170258	Oct 2019
		WS0278-GP 4x N-Glycanase® 1 ea (700 µl)	W170180	Oct 2019

### QC RELEASE

The components were assayed following the booklet protocol and conform to specifications<sup>1</sup>.

\_\_\_\_\_  
Authorized Signature

<sup>1</sup> Glycans were released from (50 µg) of human IgG, labeled and cleaned up using randomly selected GlykoPrep components. The labeled glycans were run on a high pressure liquid chromatography column and the relative abundance of major glycan peaks was comparable to approved historical results for the same protein lot.