



CERTIFICATE OF ANALYSIS

PRODUCT NAME: GLYKOPREP®-*plus* DIGESTION MODULE (96-ct)
PRODUCT CODE: GSP-RX
LOT NUMBER: 959 026
STORAGE: Store at 2 - 8°C.
PACK SIZE: 96 cartridges
EXPIRATION: October 2019; the components of this module are usable at least through this month.

COMPONENTS

| Component Name | Lot No. & Expiry | Contents | Lot No. | Expiry |
|--|---------------------|---|---------|----------|
| WS0253 Digestion (RX) Cartridges | W180105 Jul 2021 | | | |
| WS0256-GP Immobilization Reagent Set | W180147 Jan 2020 | WS0226-GP Denaturation Reagent 1 ea (35ml) | W180029 | Feb 2020 |
| | | WS0255-GP Blocking Reagent 1 ea (14 ml) | W180006 | Jan 2020 |
| WS0259-GP Digestion Reagent Set | W180152 Oct 2019 | WS0229-GP Finishing Reagent 1 ea (10 ml) | W170121 | Nov 2019 |
| | | WS0276-GP 25x Digestion Buffer 1 ea (1.3 ml) | W170258 | Oct 2019 |
| | | WS0278-GP 4x N-Glycanase® 1 ea (700 µl) | W170259 | Oct 2019 |

QC RELEASE

The components were assayed following the booklet protocol and conform to specifications¹.

Authorized Signature

¹ Glycans were released from (50 µg) of human IgG, labeled and cleaned up using randomly selected GlykoPrep components. The labeled glycans were run on a high pressure liquid chromatography column and the relative abundance of major glycan peaks was comparable to approved historical results for the same protein lot.