

CERTIFICATE OF ANALYSIS

PRODUCT NAME: GLYKOPREP®-plus DIGESTION MODULE (96-ct)
 PRODUCT CODE: GSP-RX
 LOT NUMBER: 959 028
 STORAGE: Store at 2 - 8°C.
 PACK SIZE: 96 cartridges
 EXPIRATION: June 2020; the components of this module are usable at least through this month.

COMPONENTS

Component Name	Lot No. & Expiry	Contents	Lot No.	Expiry
WS0253 Digestion (RX) Cartridges	W180105 Jul 2021			
WS0256-GP Immobilization Reagent Set	W180241 Oct 2020	WS0226-GP Denaturation Reagent 1 ea (35ml)	W180217	Oct2020
		WS0255-GP Blocking Reagent 1 ea (14 ml)	W180218	Oct 2020
WS0259-GP Digestion Reagent Set	W180220 Jun 2020	WS0229-GP Finishing Reagent 1 ea (10 ml)	W180148	Dec 2020
		WS0276-GP 25x Digestion Buffer 1 ea (1.3 ml)	W180150	Jun 2020
		WS0278-GP 4x N-Glycanase® 1 ea (700 µl)	W180151	Sep 2020

QC RELEASE

The components were assayed following the booklet protocol and conform to specifications¹.

Authorized Signature

¹ Glycans were released from (50 µg) of human IgG, labeled and cleaned up using randomly selected GlykoPrep components. The labeled glycans were run on a high pressure liquid chromatography column and the relative abundance of major glycan peaks was comparable to approved historical results for the same protein lot.