



## CERTIFICATE OF ANALYSIS

PRODUCT NAME: GLY-X™ DEGLYCOSYLATION MODULE (24-ct)  
PRODUCT CODE: GX24-100  
LOT NUMBER: 670 004  
STORAGE: Store at 2-8°C.  
PACK SIZE: 24 reactions  
EXPIRATION: March 2018; the components of this module are usable at least through this month.

### COMPONENTS

Component Name	Lot No. & Expiry	Contents	Lot No.	Expiry
WS0341 Gly-X Deglycosylation Plate	W160248 Nov 2021			
WS0342-24 Gly-X Enzyme Reagent Set (24-ct)	W170002 Mar 2018	W0343-24 Gly-X N-Glycanase (24-ct) 1 ea (30 µl)	W160214	Aug 2018
		WS0344 Gly-X Digestion Buffer 1 ea (240 µl)	W160203	Mar 2018
		WS0358 Gly-X Denaturant 1 ea (240 µl)	W160194	Mar 2018

### QC RELEASE

The components conform to specifications<sup>1</sup>.

\_\_\_\_\_  
Authorized Signature

<sup>1</sup> Glycans were released from a glycoprotein using randomly selected kit components. The glycans were derivatized using a Gly-X InstantPC™ Labeling Module, cleaned up with a Gly-X InstantPC Cleanup Module and then analyzed by LC. The results were comparable to a reference standard.