



CERTIFICATE OF ANALYSIS

PRODUCT NAME: GLY-X™ DEGLYCOSYLATION MODULE FOR INSTANTQ (96-ct)
PRODUCT CODE: GX96-300
LOT NUMBER: 683 004
STORAGE: Store at 2-8°C.
PACK SIZE: 96 reactions
EXPIRATION: August 2018; the components of this module are usable at least through this month.

COMPONENTS:

| Component Name | Quantity/Kit | Lot Number | Expiry |
|--|--------------|------------|----------|
| WS0341 Gly-X Deglycosylation Plate (96-well) | 1 each | W160248 | Nov 2021 |
| WS0343 Gly-X N-Glycanase (120 µl) | 1 each | W170181 | Oct 2019 |
| WS0344 Gly-X Digestion Buffer (240 µl) | 2 each | W170261 | Mar 2019 |
| WS0348 Gly-X Blocker (Optional) (300 µl) | 1 each | W170021 | Aug 2018 |
| WS0358 Gly-X Denaturant (240 µl) | 1 each | W170249 | Mar 2019 |

QC RELEASE:

The components conform to specifications¹.

Authorized Signature

¹ Glycans were released from a glycoprotein using randomly selected kit components. The glycans were derivatized using a Gly-X InstantQ™ Labeling Module, cleaned up with a Gly-X InstantQ Cleanup Module and then analyzed on the Gly-Q Capillary Electrophoresis Instrument and by LC. The results were comparable to a reference standard.