



## CERTIFICATE OF ANALYSIS

PRODUCT NAME: GLY-X™ INSTANTQ™ CLEANUP MODULE (96-ct)  
PRODUCT CODE: GX96-302  
LOT NUMBER: 686 005  
STORAGE: Store at Room Temperature (15 - 30°C)  
PACK SIZE: 96 reactions  
EXPIRATION: June 2022; the components of this module are usable at least through this month.

### COMPONENTS:

Component Name	Quantity/Kit	Lot Number	Expiry
WS0350 InstantQ Cleanup Plate (96-well)	1 each	W170173	Jun 2022
WS0352 InstantQ Collection Plate (96-well)	1 each	W170260	Oct 2022
Waste Tray	1 each		

### QC RELEASE:

The components conform to specifications<sup>1</sup>.

---

Authorized Signature

<sup>1</sup> Glycans were released from a glycoprotein using randomly selected kit components. The glycans were derivatized using a Gly-X InstantQ™ Labeling Module, cleaned up with a Gly-X InstantQ Cleanup Module and then analyzed on the Gly-Q Capillary Electrophoresis Instrument and by LC. The results were comparable to a reference standard.