

Certificate of Analysis

Beryllium
AA Standard

Catalog Number: IAA-004
Lot Number: R00171
Lot Issue Date: 02/26/2013
Expiration Date: 03/31/2020

This Certified Reference Material (CRM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system. The analyte concentration(s) were prepared and verified by an ISO Guide 34 / ISO 17025 accredited laboratory, and compared to calibration standards independently prepared using NIST SRM(s). The certified value and uncertainty value at the 95% confidence level for each analyte is determined gravimetrically.

Analyte	True Value	Analytical Method	NIST SRM
beryllium	10018 ± 50 µg/mL	ICP / ICP-MS	3105a

Matrix: 7% nitric acid in low TOC water (< 50 ppb)

ULTRA uses purified acids, 18 megaohm double deionized water, calibrated Class A glassware, and meticulously cleaned bottles in the manufacturing of ULTRA grade standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSS Z-540-1 and ISO 9001.

-Density of Solution (measured at 19.99°C ± 0.05°C): 1.0693 g/mL



ISO 17025:2005
Accredited
A2LA
Cert. No. 0851.01

ISO 9001:2000
Registered
TUV USA, Inc.
Cert. No. 06-1004



William J. Leary
Quality Assurance Manager