

Certificate of Analysis

Copper
AA Standard

Catalog Number: IAA-029
Lot Number: L01250
Job Number: J00011813
Lot Issue Date: 10/27/2010
Expiration Date: 11/30/2017

This Certified Reference Material (CRM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory.

The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below

Analyte	True Value	Analytical Method	NIST SRM
copper	10016 ± 50 µg/mL	ICP	3114


Matrix: 2% nitric acid in low TOC water (<50 ppb)

*Balances used in the manufacture of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001



ISO 17025:2005
Accredited
A2LA
Cert. No. 0851.01

ISO 9001:2000
Registered
TUV USA, Inc.
Cert. No. 06-1004


William J. Leahy
Quality Assurance Manager