



# Certificate of Analysis

**Lithium**  
**AA Standard**

**Catalog Number:** IAA-203  
**Lot Number:** J00482  
**Job Number:** J00008338  
**Lot Issue Date:** 05/16/2008  
**Expiration Date:** 06/30/2015

This Certified Reference Material (CRM) was manufactured and verified in accordance with ULTRA's ISO 9001:2000 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below

| Analyte | True Value |   |         | Analytical Method | NIST SRM |
|---------|------------|---|---------|-------------------|----------|
|         |            |   |         |                   |          |
| lithium | 1001       | ± | 5 µg/mL | ICP / ICP-MS      | 3129a    |


Matrix: 2% nitric acid in low TOC water (< 50 ppb)

ULTRA uses purified acids, 18 megohm double deionized water, calibrated Class A glassware & meticulously cleaned bottles in the manufacturing of ULTRAgrade standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001



ISO 17025:2005  
Accredited  
A2LA  
Cert. No. 0851.01

ISO 9001:2000  
Registered  
TUV USA, Inc.  
Cert. No. 06-1004

  
William J. Leary  
Quality Assurance Manager