Certificate of Analysis

Phosphorus
AA Standard

Catalog Number: IAA-215
Lot Number: CM-2788
Lot Issue Date: 05/20/2015
Expiration Date: 06/30/2022

This Certified Reference Material (CRM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system. The analyte concentrations were verified by our ISO 17025 accredited laboratory to be within ± 2.5%, when compared to calibration standards independently prepared using NIST SRM(s). The certified value and uncertainty value at the 95% confidence level for each analyte is determined gravimetrically.

Analyte	True Value				Analytical Method	NIST SRM
phosphorus	1002	±	5	µg/mL	ICP / ICP-MS	3139a

Matrix: 2% nitric acid in low TOC water (< 50 ppb)

ULTRA uses purified acids, 18 megohm double deionized water, calibrated Class A glassware & meticulously cleaned bottles in the manufacturing of ULTRAgrade standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.

-Density of Solution (measured at 20.00°C ± 0.05°C): 1.0102 g/mL

William J. Leary
Quality Assurance Manager

ANAB ACCREDITED ISO/IEC 17025 TESTING LABORATORY

ISO 9001 Registered Quality System – TUV USA