



Certificate of Analysis

Calcium
AA Standard

Catalog Number: IAA-220
Lot Number: J01264
Job Number: J00009241
Lot Issue Date: 01/05/2009
Expiration Date: 01/31/2016

This Certified Reference Material (CRM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system. The analyte concentrations were verified by our ISO 17025 accredited laboratory to be within $\pm 2.5\%$, when compared to calibration standards independently prepared using NIST SRM(s). The certified value and uncertainty value for each analyte is determined gravimetrically.

Analyte	True Value	Analytical Method	NIST SRM
calcium	1002 $\pm$ 3 $\mu\text{g/mL}$	ICP / ICP-MS	3109a


Matrix: 2% nitric acid in low TOC water (< 50 ppb)

ULTRA uses purified acids, 18 megohm double deionized water, calibrated Class A glassware & meticulously cleaned bottles in the manufacturing of ULTRAgrade standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001



ISO 17025:2005
Accredited
A2LA
Cert. No. 0851.01

ISO 9001:2000
Registered
TUV USA, Inc.
Cert. No. 06-1004


William J. Leary
Quality Assurance Manager