Certificate of Analysis

Strontium
AA Standard

Catalog Number: IAA-238
Lot Number: M00082
Job Number: J00012126
Lot Issue Date: 01/28/2011
Expiration Date: 02/28/2018

This Certified Reference Material (CRM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system. The analyte concentrations were verified by our ISO 17025 accredited laboratory to be within ± 2.5%, when compared to calibration standards independently prepared using NIST SRM(s). The certified value and uncertainty value at the 95% confidence level for each analyte is determined gravimetrically.

Analyte	True Value				Analytical Method	NIST SRM
strontium	1001	±	5	µg/mL	ICP / ICP-MS	3153a

Matrix: 2% nitric acid in low TOC water (< 50 ppb)

ULTRA uses purified acids, 18 megohm double deionized water, calibrated Class A glassware & meticulously cleaned bottles in the manufacturing of ULTRAgrade standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.

-Density of Solution (measured at 20.0°C ± 0.5°C): 1.016 g/mL

ACCREDITED

ISO 17025:2005
Accredited
A2LA
Cert. No. 0851.01

ISO 9001:2000
Registered
TUV USA, Inc.
Cert. No. 06-1004

William J. Leary
Quality Assurance Manager