



Certificate of Analysis

**Zirconium
AA Standard**

Catalog Number: IAA-240
Lot Number: J00470
Job Number: J00008194
Lot Issue Date: 05/14/2008
Expiration Date: 06/30/2015

This Certified Reference Material (CRM) was manufactured and verified in accordance with ULTRA's ISO 9001:2000 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below

Analyte	True Value			Analytical Method	NIST SRM
zirconium	1000	±	5	µg/mL	ICP / ICP-MS


Matrix: 1% nitric acid in low TOC water (< 50 ppb)

ULTRA uses purified acids, 18 megohm double deionized water, calibrated Class A glassware & meticulously cleaned bottles in the manufacturing of ULTRAgrade standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001



ISO 17025:2005
Accredited
A2LA
Cert. No. 0851.01

ISO 9001:2000
Registered
TUV USA, Inc.
Cert. No. 06-1004


William J. Leary
Quality Assurance Manager