Certificate of Analysis

Niobium
AA Standard

Catalog Number: IAA-241
Lot Number: L00074
Job Number: J00010826
Lot Issue Date: 01/21/2010
Expiration Date: 02/28/2017

This Certified Reference Material (CRM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below

Analyte	True Value				Analytical Method	NIST SRM
niobium	1000	±	5	µg/mL	ICP / ICP-MS	3137

Matrix: trace hydrofluoric acid in low TOC water (< 50 ppb)

ULTRA uses purified acids, 18 megohm double deionized water, calibrated Class A glassware & meticulously cleaned bottles in the manufacturing of ULTRAgrade standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001

ACCREDITED

ISO 17025:2005
Accredited
A2LA
Cert. No. 0851.01

ISO 9001:2000
Registered
TUV USA, Inc.
Cert. No. 06-1004

William J. Leary
Quality Assurance Manager