



Certificate of Analysis

Rhodium
AA Standard

Catalog Number: IAA-245
Lot Number: L01341
Job Number: J00012117
Lot Issue Date: 11/18/2010
Expiration Date: 12/31/2017

This Certified Reference Material (CRM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system. The analyte concentrations were verified by our ISO 17025 accredited laboratory to be within $\pm 2.5\%$, when compared to calibration standards independently prepared using NIST SRM(s). The certified value and uncertainty value at the 95% confidence level for each analyte is determined gravimetrically.

Analyte	True Value			Analytical Method	NIST SRM
rhodium	1000	$\pm$	5 $\mu\text{g/mL}$	ICP / ICP-MS	3144


Matrix: 5% hydrochloric acid in low TOC water (< 50 ppb)

ULTRA uses purified acids, 18 megohm double deionized water, calibrated Class A glassware & meticulously cleaned bottles in the manufacturing of ULTRAgrade standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.



ISO 17025:2005
Accredited
A2LA
Cert. No. 0851.01

ISO 9001:2000
Registered
TUV USA, Inc.
Cert. No. 06-1004


William J. Leary
Quality Assurance Manager