



IC Cations Mixture #1

Catalog Number: ICC-300

Lot Number: CS-1334

Lot Issue Date: 3/22/2018

Expiration Date: 4/30/2022


This Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system. The analyte concentration(s) were prepared and verified by an ISO Guide 34 / ISO 17025 accredited laboratory, and compared to calibration standards independently prepared using NIST SRM(s). The certified value and uncertainty value at the 95% confidence level for each analyte is determined gravimetrically.

Analyte	True Value			Analytical Method	NIST SRM
ammonium	400.7	±	2.0 µg/mL	IC	n/a
calcium	1003	±	5.0 µg/mL	IC	3109
lithium	50.1	±	0.3 µg/mL	IC	3129
magnesium	200.2	±	1.0 µg/mL	IC	3131
potassium	200.3	±	1.0 µg/mL	IC	3141
sodium	200.7	±	1.0 µg/mL	IC	3152

Matrix: 0.2% nitric acid (trace) in low TOC water (< 50 ppb)

ULTRA uses purified acids, 18 megaohm double deionized water, calibrated Class A glassware, and meticulously cleaned bottles in the manufacturing of ULTRAGrade standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.


John Russo
President


Monica Bourgeois
Director of QA/RA



ISO 9001 Registered Quality System – TUV USA