

Certificate of Analysis

IC Cations Mixture #1
Catalog Number: ICC-300
Lot Number: T00787
Job Number: J00018528
Lot Issue Date: 07/23/2014
Expiration Date: 08/31/2018

This Certified Reference Material (CRM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system. The analyte concentration(s) were prepared and verified by an ISO Guide 34 / ISO 17025 accredited laboratory, and compared to calibration standards independently prepared using NIST SRM(s). The certified value and uncertainty value at the 95% confidence level for each analyte is determined gravimetrically.

Analyte	True Value			Analytical Method	NIST SRM
ammonium	400.0	±	2.0 µg/mL	IC	n/a
calcium	1000	±	5.0 µg/mL	IC	3109a
lithium	50.00	±	0.25 µg/mL	IC	3129a
magnesium	200.0	±	1.0 µg/mL	IC	3131a
potassium	200.0	±	1.0 µg/mL	IC	3141a
sodium	200.0	±	1.0 µg/mL	IC	3152a

Matrix: 0.2% nitric acid (trace) in low TOC water (< 50 ppb)

ULTRA uses purified acids, 18 megaohm double deionized water, calibrated Class A glassware, and meticulously cleaned bottles in the manufacturing of ULTRAgade standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.



ISO 17025:2005
Accredited
A2LA
Cert. No. 0851.01

ISO 9001:2000
Registered
TUV USA, Inc.
Cert. No. 06-1004



William J. Leary
Quality Assurance Manager