



Certificate of Analysis

IC Cations Mixture #2

Catalog Number: ICC-310

Lot Number: CS-1610

Lot Issue Date: 04/03/2018

Expiration Date: 05/31/2022


This Certified Reference Material (CRM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system. The analyte concentration(s) were prepared and verified by an ISO Guide 34 / ISO 17025 accredited laboratory, and compared to calibration standards independently prepared using NIST SRM(s). The certified value and uncertainty value at the 95% confidence level for each analyte is determined gravimetrically.

Analyte	True Value			Analytical Method	NIST SRM	
ammonium	100.0	±	0.5	µg/mL	IC	n/a
lithium	10.0	±	0.1	µg/mL	IC	3129
potassium	50.1	±	0.3	µg/mL	IC	3141
sodium	100.0	±	0.5	µg/mL	IC	3152

Matrix: 0.2% nitric acid (trace) in low TOC water (< 50 ppb)

ULTRA uses purified acids, 18 megaohm double deionized water, calibrated Class A glassware, and meticulously cleaned bottles in the manufacturing of ULTRAggrade standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.


John Russo
President


Monica Bourgeois
Director of QA/RA



ISO 9001 Registered Quality System – TUV USA