

Certificate of Analysis

ICP Calibration Standard

Catalog Number: ICM-102

Lot Number: M01101

Job Number: J00013806

Lot Issue Date: 10/17/2011

Expiration Date: 11/30/2015

This Certified Reference Material (CRM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system. The analyte concentrations were verified by our ISO 17025 accredited laboratory to be within $\pm 5.0\%$, when compared to calibration standards independently prepared using NIST SRM(s). The certified value and uncertainty value at the 95% confidence level for each analyte is determined gravimetrically.

Analyte	True Value				Analytical Method	NIST SRM
aluminum	100.0	$\pm$	0.5	$\mu\text{g/mL}$	ICP / ICP-MS	3101a
barium	5.000	$\pm$	0.025	$\mu\text{g/mL}$	ICP / ICP-MS	3104a
beryllium	1.000	$\pm$	0.005	$\mu\text{g/mL}$	ICP / ICP-MS	3105a
bismuth	200.0	$\pm$	1.0	$\mu\text{g/mL}$	ICP / ICP-MS	3106
boron	15.00	$\pm$	0.08	$\mu\text{g/mL}$	ICP / ICP-MS	3107
cadmium	20.00	$\pm$	0.10	$\mu\text{g/mL}$	ICP / ICP-MS	3108
chromium	25.00	$\pm$	0.13	$\mu\text{g/mL}$	ICP / ICP-MS	3112a
cobalt	20.00	$\pm$	0.10	$\mu\text{g/mL}$	ICP / ICP-MS	3113
copper	20.00	$\pm$	0.10	$\mu\text{g/mL}$	ICP / ICP-MS	3114
gallium	150.0	$\pm$	0.8	$\mu\text{g/mL}$	ICP / ICP-MS	3119a
indium	200.0	$\pm$	1.0	$\mu\text{g/mL}$	ICP / ICP-MS	3124a
iron	15.00	$\pm$	0.08	$\mu\text{g/mL}$	ICP / ICP-MS	3126a
lead	200.0	$\pm$	1.0	$\mu\text{g/mL}$	ICP / ICP-MS	3128
manganese	5.000	$\pm$	0.025	$\mu\text{g/mL}$	ICP / ICP-MS	3132
nickel	50.00	$\pm$	0.25	$\mu\text{g/mL}$	ICP / ICP-MS	3136
* silver	50.00	$\pm$	0.25	$\mu\text{g/mL}$	ICP / ICP-MS	3151
strontium	1.000	$\pm$	0.005	$\mu\text{g/mL}$	ICP / ICP-MS	3153a
thallium	400.0	$\pm$	2.0	$\mu\text{g/mL}$	ICP / ICP-MS	3158
zinc	20.00	$\pm$	0.10	$\mu\text{g/mL}$	ICP / ICP-MS	3168a

Matrix: 5% nitric acid in low TOC water (< 50 ppb)


** light sensitive*

ULTRA uses purified acids, 18 megohm double deionized water, calibrated Class A glassware & meticulously cleaned bottles in the manufacturing of ULTRAgrade standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001



ISO 17025:2005
Accredited
A2LA
Cert. No. 0851.01

ISO 9001:2000
Registered
TUV USA, Inc.
Cert. No. 06-1004



William J. Leary
Quality Assurance Manager