

# Certificate of Analysis

**ICP Calibration Standard**

**Catalog Number: ICM-103**

**Lot Number: T00291**

**Job Number: J00017825**

**Lot Issue Date: 03/19/2014**

**Expiration Date: 04/30/2018**

This Certified Reference Material (CRM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system. The analyte concentrations were verified by our ISO 17025 accredited laboratory to be within  $\pm 5.0\%$ , when compared to calibration standards independently prepared using NIST SRM(s). The certified value and uncertainty value at the 95% confidence level for each analyte is determined gravimetrically.

Analyte	True Value				Analytical Method	NIST SRM
aluminum	1001	$\pm$	5	$\mu\text{g/mL}$	ICP / ICP-MS	3101a
barium	1002	$\pm$	5	$\mu\text{g/mL}$	ICP / ICP-MS	3104a
bismuth	1000	$\pm$	5	$\mu\text{g/mL}$	ICP / ICP-MS	3106
boron	1002	$\pm$	5	$\mu\text{g/mL}$	ICP / ICP-MS	3107
cadmium	1002	$\pm$	5	$\mu\text{g/mL}$	ICP / ICP-MS	3108
calcium	1002	$\pm$	5	$\mu\text{g/mL}$	ICP / ICP-MS	3109a
chromium	1001	$\pm$	5	$\mu\text{g/mL}$	ICP / ICP-MS	3112a
cobalt	1001	$\pm$	5	$\mu\text{g/mL}$	ICP / ICP-MS	3113
copper	1001	$\pm$	5	$\mu\text{g/mL}$	ICP / ICP-MS	3114
gallium	1000	$\pm$	5	$\mu\text{g/mL}$	ICP / ICP-MS	3119a
indium	1000	$\pm$	5	$\mu\text{g/mL}$	ICP / ICP-MS	3124a
iron	1001	$\pm$	5	$\mu\text{g/mL}$	ICP / ICP-MS	3126a
lead	1000	$\pm$	5	$\mu\text{g/mL}$	ICP / ICP-MS	3128
lithium	1002	$\pm$	5	$\mu\text{g/mL}$	ICP / ICP-MS	3129a
magnesium	1001	$\pm$	5	$\mu\text{g/mL}$	ICP / ICP-MS	3131a
manganese	1001	$\pm$	5	$\mu\text{g/mL}$	ICP / ICP-MS	3132
nickel	1001	$\pm$	5	$\mu\text{g/mL}$	ICP / ICP-MS	3136
potassium	1001	$\pm$	5	$\mu\text{g/mL}$	ICP / ICP-MS	3141a
* silver	1001	$\pm$	5	$\mu\text{g/mL}$	ICP / ICP-MS	3151
sodium	1002	$\pm$	5	$\mu\text{g/mL}$	ICP / ICP-MS	3152a
strontium	1001	$\pm$	5	$\mu\text{g/mL}$	ICP / ICP-MS	3153a
thallium	1001	$\pm$	5	$\mu\text{g/mL}$	ICP / ICP-MS	3158
zinc	1001	$\pm$	5	$\mu\text{g/mL}$	ICP / ICP-MS	3168a

Matrix: 5% nitric acid in low TOC water (< 50 ppb)

\* *light sensitive*

ULTRA uses purified acids, 18 megohm double deionized water, calibrated Class A glassware & meticulously cleaned bottles in the manufacturing of ULTRAgrade standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.



ISO 17025:2005  
Accredited  
A2LA  
Cert. No. 0851.01

ISO 9001:2000  
Registered  
TUV USA, Inc.  
Cert. No. 06-1004



William J. Leary  
Quality Assurance Manager