

Certificate of Analysis

EPA Method 200.7**Plasma Solution****Catalog Number: ICM-237****Lot Number: CL-4507****Lot Issue Date: 11/10/2014****Expiration Date: 12/31/2017**

This Certified Reference Material (CRM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system. The analyte concentration(s) were prepared and verified by an ISO Guide 34 / ISO 17025 accredited laboratory, and compared to calibration standards independently prepared using NIST SRM(s). The certified value and uncertainty value at the 95% confidence level for each analyte is determined gravimetrically.

Analyte	True Value				Analytical Method	NIST SRM
arsenic	10.0	±	0.1	µg/mL	ICP / ICP-MS	3103a
lead	10.0	±	0.1	µg/mL	ICP / ICP-MS	3128
selenium	10.0	±	0.1	µg/mL	ICP / ICP-MS	3149
thallium	10.0	±	0.1	µg/mL	ICP / ICP-MS	3158


Matrix: 2% nitric acid in low TOC water (< 50 ppb)

ULTRA uses purified acids, 18 megaohm double deionized water, calibrated Class A glassware, and meticulously cleaned bottles in the manufacturing of ULTRA grade standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.



ISO 17025:2005
Accredited
A2LA
Cert. No. 0851.01

ISO 9001:2000
Registered
TUV USA, Inc.
Cert. No. 06-1004


William J. Leary
Quality Assurance Manager