



Certificate of Analysis

EPA Method 200.7

LPC Solution A

Catalog Number: ICM-240A

Lot Number: CP-2520

Lot Issue Date: 05/23/2016

Expiration Date: 06/30/2017

This Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system. The analyte concentration(s) were prepared and verified by an ISO Guide 34 / ISO 17025 accredited laboratory, and compared to calibration standards independently prepared using NIST SRM(s). The certified value and uncertainty value at the 95% confidence level for each analyte is determined gravimetrically.


Analyte	True Value				Analytical Method	NIST SRM
aluminum	200.3	±	1.0	µg/mL	ICP / ICP-MS	3101a
arsenic	200.2	±	1.0	µg/mL	ICP / ICP-MS	3103a
barium	200.3	±	1.0	µg/mL	ICP / ICP-MS	3104a
beryllium	200.4	±	1.0	µg/mL	ICP / ICP-MS	3105a
boron	200.7	±	1.0	µg/mL	ICP / ICP-MS	3107
cadmium	200.1	±	1.0	µg/mL	ICP / ICP-MS	3108
calcium	200.3	±	1.0	µg/mL	ICP / ICP-MS	3109a
cerium	200.3	±	1.0	µg/mL	ICP / ICP-MS	3110
chromium	200.3	±	1.0	µg/mL	ICP / ICP-MS	3112a
cobalt	200.3	±	1.0	µg/mL	ICP / ICP-MS	3113
copper	200.3	±	1.0	µg/mL	ICP / ICP-MS	3114
iron	200.2	±	1.0	µg/mL	ICP / ICP-MS	3126a
lead	200.2	±	1.0	µg/mL	ICP / ICP-MS	3128
lithium	200.3	±	1.0	µg/mL	ICP / ICP-MS	3129a
magnesium	200.2	±	1.0	µg/mL	ICP / ICP-MS	3131a
manganese	200.1	±	1.0	µg/mL	ICP / ICP-MS	3132
* mercury	200.1	±	1.0	µg/mL	ICP / ICP-MS	3133
nickel	200.2	±	1.0	µg/mL	ICP / ICP-MS	3136
phosphorus	1001	±	5	µg/mL	ICP / ICP-MS	3139a
potassium	1000	±	5	µg/mL	ICP / ICP-MS	3141a
selenium	200.1	±	1.0	µg/mL	ICP / ICP-MS	3149
* silver	25.00	±	0.13	µg/mL	ICP / ICP-MS	3151
sodium	200.3	±	1.0	µg/mL	ICP / ICP-MS	3152a
strontium	200.2	±	1.0	µg/mL	ICP / ICP-MS	3153a
thallium	200.3	±	1.0	µg/mL	ICP / ICP-MS	3158
* vanadium	200.2	±	1.0	µg/mL	ICP / ICP-MS	3165
zinc	200.3	±	1.0	µg/mL	ICP / ICP-MS	3168a


Matrix: 5% nitric acid and trace hydrofluoric acid in low TOC water (< 50 ppb)

* light sensitive

ULTRA uses purified acids, 18 megohm double deionized water, calibrated Class A glassware & meticulously cleaned bottles in the manufacturing of ULTRAgrade standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.




Peter A. King, Ph.D.
VP, Technical Operations


Daniel J. Lamendola
Director of QA/RA

ISO 9001 Registered Quality System – TUV USA