



Certificate of Analysis

Primary Safe Drinking Water Act

Standard #3; Mercury

Catalog Number: ICM-303

Lot Number: CM-3672

Lot Issue Date: 07/09/2015

Expiration Date: 08/31/2018

This Certified Reference Material (CRM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system. The analyte concentration(s) were prepared and verified by an ISO Guide 34 / ISO 17025 accredited laboratory, and compared to calibration standards independently prepared using NIST SRM(s). The certified value and uncertainty value at the 95% confidence level for each analyte is determined gravimetrically.

Analyte	True Value			Analytical Method	NIST SRM
* mercury	100.1	±	0.5 µg/mL	ICP / ICP-MS	3133

Matrix: 2% nitric acid in low TOC water (< 50 ppb)

** light sensitive*

ULTRA uses purified acids, 18 megaohm double deionized water, calibrated Class A glassware, and meticulously cleaned bottles in the manufacturing of ULTRA grade standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.



ISO 9001 Registered Quality System – TUV USA


William J. Leary
Quality Assurance Manager