



Certificate of Analysis

CLP ICP Calibration Verification Standard #2
Antimony Standard at 500 ug/mL


Catalog Number: ICM-432
Lot Number: CS-1525
Lot Issue Date: 04/02/2018
Expiration Date: 04/30/2025


This Certified Reference Material (CRM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system. The analyte concentration(s) were prepared and verified by an ISO Guide 34 / ISO 17025 accredited laboratory, and compared to calibration standards independently prepared using NIST SRM(s). The certified value and uncertainty value at the 95% confidence level for each analyte is determined gravimetrically.

Analyte	True Value			Analytical Method	NIST SRM
antimony	500.0	±	2.5 µg/mL	ICP / ICP-MS	3102

Matrix: 2% nitric acid and trace tartaric acid in low TOC water (< 50 ppb)

ULTRA uses purified acids, 18 megaohm double deionized water, calibrated Class A glassware, and meticulously cleaned bottles in the manufacturing of ULTRAgrade standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.


John Russo
President


Monica Bourgeois
Director of QA/RA



ISO 9001 Registered Quality System – TUV USA