

# Certificate of Analysis

**CLP ICP Interference Check Standard #1**

**Catalog Number: ICM-441**

**Lot Number: CM-5693**

**Lot Issue Date: 11/04/2015**

**Expiration Date: 11/30/2018**

This Certified Reference Material (CRM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system. The analyte concentration(s) were prepared and verified by an ISO Guide 34 / ISO 17025 accredited laboratory, and compared to calibration standards independently prepared using NIST SRM(s). The certified value and uncertainty value at the 95% confidence level for each analyte is determined gravimetrically.

Analyte	True Value			Analytical Method	NIST SRM	
aluminum	5004	±	25	µg/mL	ICP / ICP-MS	3101a
calcium	5005	±	25	µg/mL	ICP / ICP-MS	3109a
iron	2005	±	10	µg/mL	ICP / ICP-MS	3126a
magnesium	5003	±	25	µg/mL	ICP / ICP-MS	3131a

**Matrix:** 2% nitric acid in low TOC water (< 50 ppb)

ULTRA uses purified acids, 18 megaohm double deionized water, calibrated Class A glassware, and meticulously cleaned bottles in the manufacturing of ULTRAgrade standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.