

Certificate of Analysis

EPA Method 6020A
Interference Check Standard A

Catalog Number: ICM-635
Lot Number: CL-5274
Lot Issue Date: 12/19/2014
Expiration Date: 01/31/2017

This Certified Reference Material (CRM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system. The analyte concentration(s) were prepared and verified by an ISO Guide 34 / ISO 17025 accredited laboratory and compared to calibration standards independently prepared using NIST SRM(s). The certified value and uncertainty value at the 95% confidence level for each analyte is determined gravimetrically.

Analyte	True Value			Analytical Method	NIST SRM	
aluminum	1001	±	5	µg/mL	ICP / ICP-MS	3101a
carbon	2003	±	10	µg/mL	gravimetric	ANSI/NCSL Z-540-1
calcium	3002	±	15	µg/mL	ICP / ICP-MS	3109a
chloride	21223	±	106	µg/mL	IC	3182
iron	2504	±	13	µg/mL	ICP / ICP-MS	3126a
potassium	1002	±	5	µg/mL	ICP / ICP-MS	3141a
magnesium	1000	±	5	µg/mL	ICP / ICP-MS	3131a
molybdenum	20.0	±	0.1	µg/mL	ICP / ICP-MS	3134
sodium	2504	±	13	µg/mL	ICP / ICP-MS	3152a
phosphorus	1001	±	5	µg/mL	ICP / ICP-MS	3139a
sulfur	1001	±	5	µg/mL	ICP / ICP-MS	3154
titanium	20.0	±	0.1	µg/mL	ICP / ICP-MS	3162a


Matrix: 5% nitric acid, trace hydrofluoric acid in low TOC water (< 50 ppb)

ULTRA uses purified acids, 18 megohm double deionized water, calibrated Class A glassware & meticulously cleaned bottles in the manufacturing of ULTRAgrade standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.



ISO 17025:2005
 Accredited
 A2LA
 Cert. No. 0851.01

ISO 9001:2000
 Registered
 TUV USA, Inc.
 Cert. No. 06-1004



William J. Leahy
 Quality Assurance Manager