

# Certificate of Analysis

**EPA Method 6020A**  
**Interference Check Standard A**

**Catalog Number: ICM-635**  
**Lot Number: CS-1482**  
**Lot Issue Date: 03/28/2018**  
**Expiration Date: 04/30/2020**


This Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system. The analyte concentration(s) were prepared and verified by an ISO Guide 34 / ISO 17025 accredited laboratory and compared to calibration standards independently prepared using NIST SRM(s). The certified value and uncertainty value at the 95% confidence level for each analyte is determined gravimetrically.

Analyte	True Value			Analytical Method	NIST SRM
aluminum	1001	±	5	µg/mL	ICP/ ICP-MS 3101
carbon	2001	±	10	µg/mL	gravimetric ANSI/NCSLZ-540-1
calcium	3003	±	15	µg/mL	ICP/ ICP-MS 3109
chloride	21216	±	106	µg/mL	IC 3182
iron	2501	±	13	µg/mL	ICP/ ICP-MS 3126
potassium	1000	±	5	µg/mL	ICP/ ICP-MS 3141
magnesium	1000	±	5	µg/mL	ICP/ ICP-MS 3131
molybdenum	20.0	±	0.1	µg/mL	ICP/ ICP-MS 3134
sodium	2502	±	13	µg/mL	ICP/ ICP-MS 3152
phosphorus	1000	±	5	µg/mL	ICP/ ICP-MS 3139
sulfur	1001	±	5	µg/mL	ICP/ ICP-MS 3154
titanium	20.0	±	0.1	µg/mL	ICP/ ICP-MS 3162

Matrix: 5% nitric acid, trace hydrofluoric acid in low TOC water (< 50 ppb)

ULTRA uses purified acids, 18 megohm double deionized water, calibrated Class A glassware & meticulously cleaned bottles in the manufacturing of ULTRAgrade standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.

  
 John Russo  
 President

  
 Monica Bourgeois  
 Director of QA/RA



ISO 9001 Registered Quality System – TUV USA