

Certificate of Analysis

EPA Method 200.8
Standard Solution B

Catalog Number: ICM-802
Lot Number: CL-2515
Lot Issue Date: 07/01/2014
Expiration Date: 07/31/2017

This Certified Reference Material (CRM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system. The analyte concentration(s) were prepared and verified by an ISO Guide 34 / ISO 17025 accredited laboratory, and compared to calibration standards independently prepared using NIST SRM(s). The certified value and uncertainty value at the 95% confidence level for each analyte is determined gravimetrically.

Analyte	True Value				Analytical Method	NIST SRM
barium	10.0	±	0.1	µg/mL	ICP / ICP-MS	3104a
* silver	10.0	±	0.1	µg/mL	ICP / ICP-MS	3151

Matrix: 2% nitric acid in low TOC water (< 50 ppb)

** light sensitive*

ULTRA uses purified acids, 18 megaohm double deionized water, calibrated Class A glassware, and meticulously cleaned bottles in the manufacturing of ULTRA grade standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.