

Certificate of Analysis

EPA Method 200.8
Internal Standard Solution

Catalog Number: ICM-810
Lot Number: CM-6301
Lot Issue Date: 12/16/2015
Expiration Date: 01/31/2020

This Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system. The analyte concentration(s) were prepared and verified by an ISO Guide 34 / ISO 17025 accredited laboratory, and compared to calibration standards independently prepared using NIST SRM(s). The certified value and uncertainty value at the 95% confidence level for each analyte is determined gravimetrically.

Analyte	True Value				Analytical Method	NIST SRM
scandium	100.1	±	0.5	µg/mL	ICP / ICP-MS	3148a
yttrium	100.2	±	0.5	µg/mL	ICP / ICP-MS	3167a
indium	100.0	±	0.5	µg/mL	ICP / ICP-MS	3124a
terbium	100.0	±	0.5	µg/mL	ICP / ICP-MS	3157a
bismuth	100.0	±	0.5	µg/mL	ICP / ICP-MS	3106

Matrix: 2% nitric acid in low TOC water (< 50 ppb)

ULTRA uses purified acids, 18 megaohm double deionized water, calibrated Class A glassware, and meticulously cleaned bottles in the manufacturing of ULTRA grade standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.