



Certificate of Analysis

ULTRAGrade™ Solution
Germanium ICP Standard
1000 µg/mL

Catalog Number: ICP-032
Lot Number: M01174
Job Number: J00013879
Lot Issue Date: 11/07/2011
Expiration Date: 12/31/2018

Starting Material: Germanium Dioxide
Starting Material Purity: 99.999%
Starting Material Lot No.: BH01975
Matrix: low TOC water (< 50 ppb) with trace HF
Atomic Weight Ge: 72.60

Certified Value: 1001 ± 2 µg/mL

This Certified Reference Material (CRM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system. The analyte concentrations were verified by our ISO 17025 accredited laboratory to be within ± 2.5%, when compared to calibration standards independently prepared using NIST SRM(s). The certified value and uncertainty value at the 95% confidence level for each analyte is determined gravimetrically.

Classical Wet Assay Method: Theoretical, based on gravimetric measurements

Confirmation by Inductively Coupled Plasma Spectroscopy (ICP / ICP-MS) vs. NIST SRM 3120a

ULTRA uses purified acids, 18 megohm double deionized water, calibrated Class A glassware & meticulously cleaned bottles in the manufacturing of ULTRAGrade standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z-540-1 and ISO 9001.

Trace Metallic Impurities in Solution Standard in µg/mL:

* Al	<0.005	D	* Ga	<0.005	ND	n Nb		n S
* Sb	<0.005	ND	s Ge			n Os		n Ta
* As	<0.005	ND	n Au			* Pd	<0.005	n Te
* Ba	<0.005	ND	n Hf			* P	<0.005	n Tb
* Be	<0.005	ND	n Ho			* Pt	<0.005	n Tl
* Bi	<0.005	ND	* In	<0.005	ND	* K	<0.005	n Th
* B	<0.005	ND	n Ir			n Pr		n Tm
* Cd	<0.005	ND	* Fe	<0.005	ND	n Re		* Sn
* Ca	<0.005	D	* La	<0.005	ND	n Rh		* Ti
n Ce			* Pb	<0.005	ND	n Rb		n W
n Cs			* Li	<0.005	ND	n Ru		n U
* Cr	<0.005	ND	n Lu			n Sm		* V
* Co	<0.005	ND	* Mg	<0.005	D	n Sc		n Yb
* Cu	<0.005	ND	* Mn	<0.005	ND	* Se	<0.005	n Y
n Dy			* Hg	<0.005	ND	* Si	<0.005	* Zn
* Er	<0.005	ND	* Mo	<0.005	ND	* Ag	<0.005	n Zr
* Eu	<0.005	ND	n Nd			* Na	<0.005	
* Gd	<0.005	ND	* Ni	<0.005	ND	* Sr	<0.005	ND

* - element checked for
ND - not detected

i - spectral interference
D - detected

n - not checked for
s - solution standard element

Density of Solution (measured at 21.0 ± 0.5°C): 1.009 g/mL



ISO 9001 Registered Quality System – TUV USA


William J. Leary
Quality Assurance Manager