



Certificate of Analysis

ULTRAGrade™ Solution
Iridium ICP Standard
1000 µg/mL

Catalog Number: ICP-077
Lot Number: P00993
Job Number: J00015248
Lot Issue Date: 09/28/2012
Expiration Date: 10/31/2019

Starting Material: iridium trichloride
Starting Material Purity: 99.95%
Starting Material Lot No.: BH01955
Matrix: 5% hydrochloric acid
Atomic Weight Ir: 192.2

Certified Value: 1001 ± 2 µg/mL

This Certified Reference Material (CRM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system. The analyte concentrations were verified by our ISO 17025 accredited laboratory to be within ± 2.5%, when compared to calibration standards independently prepared using second source CRM(s). The certified value and uncertainty value at the 95% confidence level for each analyte is determined gravimetrically.

Classical Wet Assay Method: Theoretical, based on gravimetric measurements

Confirmation by Inductively Coupled Plasma Spectroscopy (ICP / ICP-MS)

ULTRA uses purified acids, 18 megohm double deionized water, calibrated Class A glassware & meticulously cleaned bottles in the manufacturing of ULTRAGrade standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z-540-1 and ISO 9001.

Trace Metallic Impurities in Solution Standard in µg/mL:

| | | | |
|----------------|----------------|----------------|----------------|
| * Al <0.005 ND | * Ga <0.005 ND | n Nb | n S |
| * Sb <0.005 ND | n Ge | n Os | n Ta |
| * As <0.005 ND | * Au <0.005 D | * Pd <0.005 D | n Te |
| * Ba <0.005 ND | n Hf | * P <0.005 ND | n Tb |
| * Be <0.005 ND | n Ho | * Pt <0.110 D | * Tl <0.005 ND |
| * Bi <0.005 ND | * In <0.005 ND | * K <0.008 D | n Th |
| * B <0.005 ND | s Ir | n Pr | n Tm |
| * Cd <0.005 D | * Fe <0.005 D | n Re | * Sn <0.005 ND |
| * Ca <0.010 D | * La <0.005 ND | * Rh <0.220 D | * Ti <0.005 ND |
| n Ce | * Pb <0.055 D | n Rb | n W |
| n Cs | * Li <0.005 ND | * Ru <0.005 D | n U |
| * Cr <0.005 D | n Lu | n Sm | * V <0.005 ND |
| * Co <0.005 D | * Mg <0.005 ND | n Sc | n Yb |
| * Cu <0.005 ND | * Mn <0.005 ND | * Se <0.005 ND | n Y |
| n Dy | * Hg <0.005 ND | * Si <0.005 ND | * Zn <0.005 ND |
| * Er <0.005 ND | * Mo <0.005 ND | * Ag <0.005 ND | n Zr |
| * Eu <0.005 ND | n Nd | * Na <0.046 D | |
| * Gd <0.005 ND | * Ni <0.005 ND | * Sr <0.005 ND | |

* - element checked for
ND - not detected

i - spectral interference
D - detected

n - not checked for
s - solution standard element

Density of Solution (measured at 20.00°C ± 0.05°C): 1.0104 g/mL




William J. Leary
Quality Assurance Manager

ISO 9001 Registered Quality System – TUV USA