



# Certificate of Analysis

**ULTRAGrade™ Solution**  
**Aluminum ICP / ICP-MS Standard**  
**10000 µg/mL**

**Catalog Number:** ICP-113  
**Lot Number:** CM-1334  
**Lot Issue Date:** 03/23/2015  
**Expiration Date:** 04/30/2022

**Starting Material:** aluminum nitrate nonahydrate  
**Starting Material Purity:** 99.999%  
**Starting Material Lot #:** RM07806  
**Matrix:** 2% nitric acid in low TOC water (< 50 ppb)  
**Atomic Weight Al:** 26.98

**Certified Value:** 10013 ± 20 µg/mL

This Certified Reference Material (CRM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system. The analyte concentration(s) were prepared and verified by an ISO Guide 34 / ISO 17025 accredited laboratory, and compared to calibration standards independently prepared using NIST SRM(s). The certified value and uncertainty value at the 95% confidence level for each analyte is determined gravimetrically.

**Classical Wet Assay Method:** Theoretical, based on gravimetric measurements

**Confirmation by Inductively Coupled Plasma Spectroscopy (ICP / ICP/MS) vs. NIST SRM 3101a**

ULTRA uses purified acids, 18 megohm double deionized water, calibrated Class A glassware & meticulously cleaned bottles in the manufacturing of ULTRAGrade standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.

**Trace Metallic Impurities in Solution Standard in µg/mL:**

|   |    |        |    |   |    |        |    |   |    |        |    |   |    |        |    |
|---|----|--------|----|---|----|--------|----|---|----|--------|----|---|----|--------|----|
| s | Al |        |    | * | Ga | <0.005 | D  | n | Nb |        |    | n | S  |        |    |
| * | Sb | <0.005 | ND | n | Ge |        |    | n | Os |        |    | n | Ta |        |    |
| * | As | <0.005 | ND | n | Au |        |    | * | Pd | <0.005 | ND | n | Te |        |    |
| * | Ba | <0.005 | ND | n | Hf |        |    | * | P  | <0.005 | ND | n | Tb |        |    |
| * | Be | <0.005 | ND | n | Ho |        |    | * | Pt | <0.005 | ND | * | Tl | <0.005 | ND |
| * | Bi | <0.005 | ND | * | In | <0.005 | ND | * | K  | <0.005 | ND | n | Th |        |    |
| * | B  | <0.005 | ND | n | Ir |        |    | n | Pr |        |    | n | Tm |        |    |
| * | Cd | <0.005 | ND | * | Fe | <0.005 | D  | n | Re |        |    | * | Sn | <0.005 | ND |
| * | Ca | <0.005 | ND | * | La | <0.005 | ND | n | Rh |        |    | * | Ti | <0.005 | ND |
| n | Ce |        |    | * | Pb | <0.005 | ND | n | Rb |        |    | n | W  |        |    |
| n | Cs |        |    | * | Li | <0.005 | ND | n | Ru |        |    | n | U  |        |    |
| * | Cr | <0.005 | ND | n | Lu |        |    | n | Sm |        |    | * | V  | <0.005 | ND |
| * | Co | <0.005 | ND | * | Mg | <0.005 | ND | n | Sc |        |    | n | Yb |        |    |
| * | Cu | <0.005 | ND | * | Mn | <0.005 | ND | * | Se | <0.005 | ND | n | Y  |        |    |
| n | Dy |        |    | * | Hg | <0.005 | ND | * | Si | <0.005 | ND | * | Zn | <0.005 | ND |
| * | Er | <0.005 | ND | * | Mo | <0.005 | ND | * | Ag | <0.005 | ND | n | Zr |        |    |
| * | Eu | <0.005 | ND | n | Nd |        |    | * | Na | <0.005 | D  |   |    |        |    |
| * | Gd | <0.005 | ND | * | Ni | <0.005 | ND | * | Sr | <0.005 | ND |   |    |        |    |

\* - element checked for  
ND - not detected

I - spectral interference  
D - detected

n - not checked for  
s - solution standard element

**Density of Solution (measured at 20.00°C ± 0.05°C):** 1.0682 g/mL



ISO 17025:2005  
Accredited  
A2LA  
Cert. No. 0851.01

ISO 9001:2000  
Registered  
TUV USA, Inc.  
Cert. No. 06-1004

William J. Leary  
Quality Assurance Manager