



Certificate of Analysis

ULTRAGrade™ Solution
Calcium ICP / ICP-MS Standard
10000 µg/mL

Catalog Number: ICP-120
Lot Number: CM-4634
Lot Issue Date: 09/01/2015
Expiration Date: 09/30/2022

Starting Material: calcium carbonate
Starting Material Purity: 99.999%
Starting Material Lot #: RM07809
Matrix: 5% nitric acid in low TOC water (< 50 ppb)
Atomic Weight Ca: 40.08

Certified Value: 10008 ± 20 µg/mL

This Certified Reference Material (CRM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system. The analyte concentration(s) were prepared and verified by an ISO Guide 34 / ISO 17025 accredited laboratory, and compared to calibration standards independently prepared using NIST SRM(s). The certified value and uncertainty value at the 95% confidence level for each analyte is determined gravimetrically.

Classical Wet Assay Method: Theoretical, based on gravimetric measurements

Confirmation by Inductively Coupled Plasma Spectroscopy (ICP / ICP/MS) vs. NIST SRM 3109a

ULTRA uses purified acids, 18 megohm double deionized water, calibrated Class A glassware & meticulously cleaned bottles in the manufacturing of ULTRAGrade standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.

Trace Metallic Impurities in Solution Standard in µg/mL:

* Al <0.005 ND	* Ga <0.005 ND	n Nb	n S
* Sb <0.005 ND	n Ge	n Os	n Ta
* As <0.005 ND	n Au	* Pd <0.005 ND	n Te
* Ba <0.005 ND	n Hf	* P <0.005 ND	n Tl <0.005 ND
* Be <0.005 ND	n Ho	* Pt <0.005 ND	n Th
* Bi <0.005 ND	* In <0.005 ND	* K <0.005 ND	n Tm
* B <0.005 ND	n Ir	n Pr	* Sn <0.005 ND
* Cd <0.005 ND	* Fe <0.005 ND	n Re	* Ti <0.005 ND
s Ca	* La <0.005 ND	n Rh	n W
n Ce	* Pb <0.005 ND	n Rb	n U
n Cs	* Li <0.005 ND	n Ru	* V <0.005 ND
* Cr <0.005 ND	n Lu	n Sm	n Yb
* Co <0.005 ND	* Mg <0.005 D	n Sc	n Y
* Cu <0.005 ND	* Mn <0.005 ND	* Se <0.005 ND	* Zn <0.005 ND
n Dy	* Hg <0.005 ND	* Si <0.005 ND	n Zr
* Er <0.005 ND	* Mo <0.005 ND	* Ag <0.005 ND	
* Eu <0.005 ND	n Nd	* Na <0.005 D	
* Gd <0.005 ND	* Ni <0.005 ND	* Sr <0.005 D	

* - element checked for
ND - not detected

l - spectral interference
D - detected

n - not checked for
s - solution standard element

Density of Solution (measured at 20.00°C ± 0.05°C): 1.0374 g/mL



ISO 9001 Registered Quality System - TUV USA

William J. Leary
Quality Assurance Manager