



ULTRAGrade™ Solution
Iron ICP / ICP-MS Standard
10000 µg/mL

Catalog Number: ICP-126
Lot Number: R00990
Job Number: J00016995
Lot Issue Date: 09/17/2013
Expiration Date: 10/30/2020

Starting Material: iron (III) nitrate nonahydrate
 Starting Material Purity: 99.9995%
 Starting Material Lot #: BH01780
 Matrix: 2% nitric acid in low TOC water (< 50 ppb)
 Atomic Weight Fe: 55.85

Certified Value: 10001 ± 20 µg/mL

This Certified Reference Material (CRM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system. The analyte concentration(s) were prepared and verified by an ISO Guide 34 / ISO 17025 accredited laboratory, and compared to calibration standards independently prepared using NIST SRM(s). The certified value and uncertainty value at the 95% confidence level for each analyte is determined gravimetrically.

Classical Wet Assay Method: Theoretical, based on gravimetric measurements

Confirmation by Inductively Coupled Plasma Spectroscopy (ICP / ICP/MS) vs. NIST SRM 3126a

ULTRA uses purified acids, 18 megohm double deionized water, calibrated Class A glassware & meticulously cleaned bottles in the manufacturing of ULTRAGrade standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.

Trace Metallic Impurities in Solution Standard in µg/mL:

* Al <0.005 D	* Ga <0.005 D	* Nb <0.005 ND	n S
* Sb <0.005 ND	n Ge	n Os	* Ta <0.005 ND
* As <0.005 ND	* Au <0.005 ND	* Pd <0.005 ND	n Te
* Ba <0.005 D	* Hf <0.005 ND	n P	* Tb <0.005 ND
* Be <0.005 ND	* Ho <0.005 ND	* Pt <0.005 ND	n Tl
n Bi	* In <0.005 D	* K <0.005 ND	* Th <0.005 ND
* B <0.005 ND	n Ir	* Pr <0.005 ND	* Tm <0.005 ND
* Cd <0.005 D	s Fe	* Re <0.005 ND	* Sn <0.005 ND
* Ca <0.005 D	* La <0.005 ND	* Rh <0.005 ND	* Ti <0.005 ND
n Ce	* Pb <0.005 ND	n Rb	* W <0.005 ND
n Cs	* Li <0.005 ND	* Ru <0.005 ND	* U <0.005 ND
* Cr <0.005 D	n Lu	* Sm <0.005 ND	* V <0.005 D
* Co <0.005 ND	* Mg <0.005 D	* Sc <0.005 ND	* Yb <0.005 ND
* Cu <0.005 D	* Mn <0.005 D	n Se	* Y <0.005 ND
* Dy <0.005 ND	n Hg	n Si	* Zn <0.005 D
* Er <0.005 ND	* Mo <0.005 ND	* Ag <0.005 ND	* Zr <0.005 ND
* Eu <0.005 ND	* Nd <0.005 ND	* Na <0.005 D	
* Gd <0.005 ND	* Ni <0.005 D	* Sr <0.005 D	

* - element checked for
 ND - not detected

l - spectral interference
 D - detected

n - not checked for
 s - solution standard element

Density of Solution (measured at 20.00°C ± 0.05°C): 1.0422 g/mL