Certificate of Analysis

ULTRAgrade™ Solution
Zirconium ICP Standard
10000 µg/mL

Catalog Number: ICP-140
Lot Number: K00141
Job Number: J00009482
Lot Issue Date: 02/16/2009
Expiration Date: 02/28/2016

Starting Material:	Zirconyl Nitrate Hydrate
Starting Material Purity:	99.99%
Starting Material Lot No.:	BH01193
Matrix:	5% nitric acid in low TOC water (< 50 ppb)
Atomic Weight Zr:	91.22

Certified Value: 10003 ± 20 µg/mL

This Certified Reference Material (CRM) was manufactured and verified in accordance with ULTRA's ISO 9001:2000 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The certified value and uncertainty value, at the 95% confidence level, for each analyte determined gravimetrically.

Classical Wet Assay Method: Theoretical, based on gravimetric measurements

Confirmation by Inductively Coupled Plasma Spectroscopy (ICP / ICP-MS) vs. NIST SRM 3169

ULTRA uses purified acids, 18 megohm double deionized water, calibrated Class A glassware & meticulously cleaned bottles in the manufacturing of ULTRAgrade standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001

Trace Metallic Impurities in Solution Standard in µg/mL:

*	Al	<0.150	D	*	Ga	<0.005	ND	n	Nb			n	S		
*	Sb	<0.005	ND	n	Ge			n	Os			n	Ta		
*	As	<0.005	ND	n	Au			*	Pd	<0.005	ND	n	Te		
*	Ba	<0.005	ND	*	Hf	<0.750	D	*	P	<0.005	ND	n	Tb		
*	Be	<0.005	ND	n	Ho			*	Pt	<0.005	ND	*	Tl	<0.005	ND
*	Bi	<0.005	ND	*	In	<0.005	ND	*	K	<0.005	ND	n	Th		
*	B	<0.005	ND	n	Ir			n	Pr			n	Tm		
*	Cd	<0.005	ND	*	Fe	<0.100	D	n	Re			*	Sn	<0.005	ND
*	Ca	<0.050	D	*	La	<0.005	ND	n	Rh			*	Ti	<0.005	ND
n	Ce			*	Pb	<0.005	ND	n	Rb			n	W		
n	Cs			*	Li	<0.005	ND	n	Ru			n	U		
*	Cr	<0.005	ND	n	Lu			n	Sm			*	V	<0.005	ND
*	Co	<0.005	ND	*	Mg	<0.010	D	n	Sc			n	Yb		
*	Cu	<0.005	ND	*	Mn	<0.005	ND	*	Se	<0.005	ND	n	Y		
n	Dy			*	Hg	<0.005	ND	*	Si	<0.005	ND	*	Zn	<0.005	ND
*	Er	<0.005	ND	*	Mo	<0.005	ND	*	Ag	<0.005	ND	s	Zr		
*	Eu	<0.005	ND	n	Nd			*	Na	<0.005	ND				
*	Gd	<0.005	ND	*	Ni	<0.005	ND	*	Sr	<0.005	ND				

* - element checked for
ND - not detected

i - spectral interference
D - detected

n - not checked for
s - solution standard element

Density of Solution (measured at 21.8°C ± 0.5): 1.041 g/mL

ANAB ACCREDITED ISO/IEC 17025 TESTING LABORATORY

William J. Leary
Quality Assurance Manager

ISO 9001 Registered Quality System – TUV USA