



Certificate of Analysis

ULTRAGrade™ Solution
Tin ICP Standard
10000 µg/mL

Catalog Number: ICP-150
Lot Number: K00799
Job Number: J00010213
Lot Issue Date: 08/06/2009
Expiration Date: 09/30/2016

Starting Material: Ammonium Hexafluorostannate
Starting Material Purity: 99.999%
Starting Material Lot No.: BH00736
Matrix: 2% nitric acid, with trace hydrofluoric acid, in low TOC water (< 50 ppb)
Atomic Weight Sn: 118.70

Certified Value: 10012 ± 20 µg/mL

This Certified Reference Material (CRM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system. The analyte concentrations were verified by our ISO 17025 accredited laboratory to be within ± 2.5%, when compared to calibration standards independently prepared using NIST SRM(s). The certified value and uncertainty value for each analyte is determined gravimetrically.

Classical Wet Assay Method: Theoretical, based on gravimetric measurements

Confirmation by Inductively Coupled Plasma Spectroscopy (ICP / ICP-MS) vs. NIST SRM 3161a

ULTRA uses purified acids, 18 megohm double deionized water, calibrated Class A glassware & meticulously cleaned bottles in the manufacturing of ULTRAGrade standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.

Trace Metallic Impurities in Solution Standard in µg/mL:

| | | | |
|----------------|----------------|----------------|----------------|
| * Al <0.005 ND | * Ga <0.005 ND | n Nb | n S |
| * Sb <0.005 ND | n Ge | n Os | n Ta |
| * As <0.005 ND | n Au | * Pd <0.005 ND | n Te |
| * Ba <0.005 ND | n Hf | * P <0.005 ND | n Tb |
| * Be <0.005 ND | n Ho | * Pt <0.005 ND | * Tl <0.005 ND |
| * Bi <0.005 ND | * In <0.005 ND | * K <0.005 ND | n Th |
| * B <0.005 ND | n Ir | n Pr | n Tm |
| * Cd <0.005 ND | * Fe <0.005 ND | n Re | s Sn |
| * Ca <0.010 D | * La <0.005 ND | n Rh | * Ti <0.005 ND |
| n Ce | * Pb <0.005 ND | n Rb | n W |
| n Cs | * Li <0.005 ND | n Ru | n U |
| * Cr <0.005 ND | n Lu | n Sm | * V <0.005 ND |
| * Co <0.005 ND | * Mg <0.010 D | n Sc | n Yb |
| * Cu <0.005 ND | * Mn <0.005 ND | * Se <0.005 ND | n Y |
| n Dy | * Hg <0.005 ND | * Si <0.005 ND | * Zn <0.005 ND |
| * Er <0.005 ND | * Mo <0.005 ND | * Ag <0.010 D | n Zr |
| * Eu <0.005 ND | n Nd | * Na <0.005 ND | |
| * Gd <0.005 ND | * Ni <0.005 ND | * Sr <0.005 ND | |

* - element checked for
ND - not detected

i - spectral interference
D - detected

n - not checked for
s - solution standard element

Density of Solution (measured at 24.0°C ± 0.5°C): 1.036 g/mL



ISO 9001 Registered Quality System – TUV USA

William J. Leary
Quality Assurance Manager