



Certificate of Analysis

ULTRAGrade™ Solution
Barium ICP Standard
10000 µg/mL

Catalog Number: ICP-156
Lot Number: K00111
Job Number: J00009433
Lot Issue Date: 02/10/2009
Expiration Date: 02/28/2016

Starting Material: Barium Nitrate
Starting Material Purity: 99.999%
Starting Material Lot No.: BH00877
Matrix: 2% nitric acid in low TOC water (< 50 ppb)
Atomic Weight Ba: 137.36

Certified Value: 10016 ± 20 µg/mL

This Certified Reference Material (CRM) was manufactured and verified in accordance with ULTRA's ISO 9001:2000 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The certified value and uncertainty value, at the 95% confidence level, for each analyte determined gravimetrically.

Classical Wet Assay Method: Theoretical, based on gravimetric measurements

Confirmation by Inductively Coupled Plasma Spectroscopy (ICP / ICP-MS) vs. NIST SRM 3104a

ULTRA uses purified acids, 18 megohm double deionized water, calibrated Class A glassware & meticulously cleaned bottles in the manufacturing of ULTRAGrade standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z-540-1 and ISO 9001

Trace Metallic Impurities in Solution Standard in µg/mL:

* Al	<0.005	ND	* Ga	<0.005	ND	n Nb		n S
* Sb	<0.005	ND	n Ge			n Os		n Ta
* As	<0.005	ND	n Au			n Pd	<0.005	n Te
s Ba			n Hf			* P	<0.005	n Tb
* Be	<0.005	ND	n Ho			* Pt	<0.005	* Tl
* Bi	<0.005	ND	* In	<0.005	ND	* K	<0.005	ND
* B	<0.005	ND	n Ir			n Pr		n Th
* Cd	<0.005	ND	* Fe	<0.005	ND	n Re		n Tm
* Ca	<0.005	ND	* La	<0.005	ND	n Rh		* Sn
n Ce			* Pb	<0.005	ND	n Rb		* Ti
n Cs			* Li	<0.005	ND	n Ru		n W
* Cr	<0.005	ND	n Lu			n Sm		n U
* Co	<0.005	ND	* Mg	<0.005	ND	n Sc		* V
* Cu	<0.005	ND	* Mn	<0.005	ND	* Se	<0.005	ND
n Dy			* Hg	<0.005	ND	* Si	<0.005	n Yb
* Er	<0.005	ND	* Mo	<0.005	ND	* Ag	<0.005	n Y
* Eu	<0.005	ND	n Nd			* Na	<0.005	* Zn
* Gd	<0.005	ND	* Ni	<0.005	ND	* Sr	<0.010	ND
								n Zr

* - element checked for
ND - not detected

i - spectral interference
D - detected

n - not checked for
s - solution standard element

Density of Solution (measured at 22.0°C ± 0.5°C): 1.027 g/mL



ISO 9001 Registered Quality System – TUV USA

William J. Leary
Quality Assurance Manager