

# Certificate of Analysis

## Custom Standard

**Product Number** ICUS-1250

**Page:** 1 of 1

**Lot Number:** CM-2635

**Lot Issue Date:** 14-May-2015

**Expiration Date:** 30-Jun-2016

This certified Reference Material (RM) was manufactured and verified in accordance with ULTRA Scientific's ISO 9001 registered quality system. A review of the gravimetric preparation data by our ISO 17025 accredited laboratory serves to verify the concentration of each analyte. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
zinc nitrate (AA grade) (as zinc)	010196-18-6	RM08153	15.0 ± 0.1 µg/mL

**Matrix:** 5% nitric acid in water

**Storage:** Store at Room Temperature (15-30°C).

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NC SL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.