



Custom Standard

Catalog Number: ICUS-3275

Lot Number: CS-1387

Lot Issue Date: 03/27/2018

Expiration Date: 04/30/2019

This Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system. The analyte concentration(s) were prepared and verified by an ISO Guide 34 / ISO 17025 accredited laboratory, and compared to calibration standards independently prepared using NIST SRM(s). The certified value and uncertainty value at the 95% confidence level for each analyte is determined gravimetrically.

Analyte	True Value			Analytical Method		NIST SRM
aluminum	100.1	±	0.5	µg/mL	ICP / ICP-MS	3101
antimony	100.0	±	0.5	µg/mL	ICP / ICP-MS	3102
arsenic	100.1	±	0.5	µg/mL	ICP / ICP-MS	3103
barium	100.1	±	0.5	µg/mL	ICP / ICP-MS	3104
beryllium	100.2	±	0.5	µg/mL	ICP / ICP-MS	3105
boron	100.1	±	0.5	µg/mL	ICP / ICP-MS	3107
cadmium	100.2	±	0.5	µg/mL	ICP / ICP-MS	3108
calcium	100.1	±	0.5	µg/mL	ICP / ICP-MS	3109
chromium	100.0	±	0.5	µg/mL	ICP / ICP-MS	3112
cobalt	100.1	±	0.5	µg/mL	ICP / ICP-MS	3113
copper	100.1	±	0.5	µg/mL	ICP / ICP-MS	3114
iron	100.2	±	0.5	µg/mL	ICP / ICP-MS	3126
lead	100.2	±	0.5	µg/mL	ICP / ICP-MS	3128
magnesium	100.0	±	0.5	µg/mL	ICP / ICP-MS	3131
manganese	100.1	±	0.5	µg/mL	ICP / ICP-MS	3132
molybdenum	100.1	±	0.5	µg/mL	ICP / ICP-MS	3134
nickel	100.2	±	0.5	µg/mL	ICP / ICP-MS	3136
potassium	100.1	±	0.5	µg/mL	ICP / ICP-MS	3141
selenium	100.2	±	0.5	µg/mL	ICP / ICP-MS	3149
silicon	100.1	±	0.5	µg/mL	ICP / ICP-MS	3150
* silver	100.1	±	0.5	µg/mL	ICP / ICP-MS	3151
sodium	100.1	±	0.5	µg/mL	ICP / ICP-MS	3152
thallium	100.1	±	0.5	µg/mL	ICP / ICP-MS	3158
titanium	100.0	±	0.5	µg/mL	ICP / ICP-MS	3162
* vanadium	100.1	±	0.5	µg/mL	ICP / ICP-MS	3165
zinc	100.2	±	0.5	µg/mL	ICP / ICP-MS	3168
phosphorus	100.2	±	0.5	µg/mL	ICP / ICP-MS	3139

Matrix: 5% nitric acid with trace hydrofluoric and tartaric acids in low TOC water (< 50 ppb)

* *light sensitive*

ULTRA uses purified acids, 18 megaohm double deionized water, calibrated Class A glassware, and meticulously cleaned bottles in the manufacturing of ULTRAgrade standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.




John Russo
President


Monica Bourgeois
Director of QA/RA

ISO 9001 Registered Quality System – TUV USA