

# Certificate of Analysis

**Custom Standard**
**Catalog Number: ICUS-4479**
**Lot Number: CM-5026**
**Lot Issue Date: 09/21/2015**
**Expiration Date: 04/30/2016**

This Certified Reference Material (CRM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system. The analyte concentration(s) were prepared and verified by an ISO Guide 34 / ISO 17025 accredited laboratory, and compared to calibration standards independently prepared using NIST SRM(s). The certified value and uncertainty value at the 95% confidence level for each analyte is determined gravimetrically.

Analyte	True Value			Analytical Method	NIST SRM	
chloride	1002	±	5.0	µg/mL	IC	3182
fluoride	1004	±	5.0	µg/mL	IC	3183
nitrate	1002	±	5.0	µg/mL	IC	3185
sulfate	1001	±	5.0	µg/mL	IC	3181
phosphate	1001	±	5.0	µg/mL	IC	3186
bromide	1001	±	5.0	µg/mL	IC	3184
nitrite	1005	±	5.0	µg/mL	IC	second source

low TOC water (< 50 ppb)

ULTRA uses purified acids, 18 megaohm double deionized water, calibrated Class A glassware, and meticulously cleaned bottles in the manufacturing of ULTRAgrade standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001.