



# Certificate of Analysis ISO Guide 34

## Custom Standard

**Product Number:** ICUS-5200-3

**Page:** 1 of 1

**Lot Number:** CS-4832

**Lot Issue Date:** 03-Oct-2018

**Expiration Date:** 30-Nov-2020

This ISO Guide 34 Reference Material (RM) was manufactured and verified in accordance with Agilent Technologies ISO 9001 registered quality system. A review of the gravimetric preparation data by our ISO 17025 accredited laboratory serves to verify the concentration of each analyte. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
beryllium acetate ( ICP grade ) (as beryllium)	000543-81-7	RM10258	40.0 ± 0.2 µg/L

**Matrix:** 1% ammonium bifluoride and trace nitric acid in water

**Storage:** Store at Room Temperature (15° to 30°C).

Agilent uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.

  
Monica Bourgeois  
QMS Representative



ISO Guide 34 Cert No.  
AR-1936

Produced in accordance with TUV USA Inc 56 100 18560026  
registered ISO 9001 Quality Management System



ISO17025 Cert No.  
AT-1937