



# Certificate of Analysis

ICP/MS Calibration Standard #2

Catalog Number: IMS-102

Lot Number: CM-5169

Lot Issue Date: 09/30/2015

Expiration Date: 10/31/2019

This Certified Reference Material (CRM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system. The analyte concentration(s) were prepared and verified by an ISO Guide 34 / ISO 17025 accredited laboratory, and compared to calibration standards independently prepared using NIST SRM(s). The certified value and uncertainty value at the 95% confidence level for each analyte is determined gravimetrically.

Analyte	True Value			Analytical Method	NIST SRM	
aluminum	10.00	±	0.05	µg/mL	ICP / ICP-MS	3101a
arsenic	10.00	±	0.05	µg/mL	ICP / ICP-MS	3103a
barium	10.00	±	0.05	µg/mL	ICP / ICP-MS	3104a
beryllium	10.00	±	0.05	µg/mL	ICP / ICP-MS	3105a
bismuth	10.00	±	0.05	µg/mL	ICP / ICP-MS	3106
cadmium	10.00	±	0.05	µg/mL	ICP / ICP-MS	3108
calcium	10.00	±	0.05	µg/mL	ICP / ICP-MS	3109a
cesium	10.00	±	0.05	µg/mL	ICP / ICP-MS	3111a
chromium	10.00	±	0.05	µg/mL	ICP / ICP-MS	3112a
cobalt	10.00	±	0.05	µg/mL	ICP / ICP-MS	3113
copper	10.00	±	0.05	µg/mL	ICP / ICP-MS	3114
gallium	10.00	±	0.05	µg/mL	ICP / ICP-MS	3119a
indium	10.00	±	0.05	µg/mL	ICP / ICP-MS	3124a
iron	10.00	±	0.05	µg/mL	ICP / ICP-MS	3126a
lead	10.00	±	0.05	µg/mL	ICP / ICP-MS	3128
lithium	10.00	±	0.05	µg/mL	ICP / ICP-MS	3129a
magnesium	10.00	±	0.05	µg/mL	ICP / ICP-MS	3131a
manganese	10.00	±	0.05	µg/mL	ICP / ICP-MS	3132
nickel	10.00	±	0.05	µg/mL	ICP / ICP-MS	3136
potassium	10.00	±	0.05	µg/mL	ICP / ICP-MS	3141a
rubidium	10.00	±	0.05	µg/mL	ICP / ICP-MS	3145a
selenium	10.00	±	0.05	µg/mL	ICP / ICP-MS	3149
* silver	10.00	±	0.05	µg/mL	ICP / ICP-MS	3151
sodium	10.00	±	0.05	µg/mL	ICP / ICP-MS	3152a
strontium	10.00	±	0.05	µg/mL	ICP / ICP-MS	3153a
thallium	10.00	±	0.05	µg/mL	ICP / ICP-MS	3158
uranium	10.00	±	0.05	µg/mL	ICP / ICP-MS	second source
* vanadium	10.00	±	0.05	µg/mL	ICP / ICP-MS	3165
zinc	10.00	±	0.05	µg/mL	ICP / ICP-MS	3168a

Matrix: 5% nitric acid, with trace hydrofluoric acid in low TOC water (< 50 ppb)

\* light sensitive

ULTRA uses purified acids, 18 megohm double deionized water, calibrated Class A glassware & meticulously cleaned bottles in the manufacturing of ULTRAgrade standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001



ISO 9001 Registered Quality System – TUV USA



William J. Leary  
Quality Assurance Manager