

Reference Material Certificate
Product Information Sheet

Product Name: ICP-MS Calibration Standard

Lot Number: 0006737883

Product Number: IMS-103

Lot Issue Date: 03-Apr-2023

Storage Conditions: Store at Room Temperature (15° to 30°C). Light Sensitive.

Expiration Date: 30-Apr-2024

Component Name	Concentration	Uncertainty	CAS#	Analyte Lot
antimony (ICP grade)	10.0 ±	0.1 µg/mL	007440-36-0	RM19157
gold (ICP grade)	10.0 ±	0.1 µg/mL	007440-57-5	RM18508
hafnium oxide (ICP grade) (as hafnium)	10.0 ±	0.1 µg/mL	012055-23-1	RM18756
iridium (III) chloride (ICP grade) (as iridium)	10.0 ±	0.1 µg/mL	010025-83-9	RM14763
palladium (ICP grade)	10.0 ±	0.1 µg/mL	007440-05-3	RM19307
platinum (ICP grade)	10.0 ±	0.1 µg/mL	007440-06-4	RM19219
rhodium (III) chloride hydrate (ICP grade) (as rhodium)	10.0 ±	0.1 µg/mL	010049-07-7	RM20112
ruthenium (III) trichloride trihydrate (ICP grade) (as ruthenium)	10.0 ±	0.1 µg/mL	007440-18-8	RM09654
tellurium (ICP grade)	10.0 ±	0.1 µg/mL	013494-80-9	RM15358
ammonium hexafluorostannate (IV) (ICP grade) (as tin)	10.0 ±	0.1 µg/mL	016919-24-7	RM11231

Matrix: water with 10% hydrochloric acid and trace tartaric, nitric, and hydrofluoric acids

Description:

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material (RM) was manufactured and verified in accordance with an ISO 9001 registered quality system. The analyte concentration(s) were prepared and verified by an ISO 17034 / ISO 17025 accredited laboratory and compared to calibration standards independently prepared using NIST SRM(s) when available. The certified value and uncertainty value at the 95% confidence level for each analyte is determined gravimetrically.

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This analytical reference standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:

This analytical reference standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.


Expiration of Certification:

The certification of this analytical reference standard is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:



Monica Bourgeois
QMS Representative