

Certificate of Analysis

ICP/MS Calibration Standard #4

Catalog Number: IMS-104

Lot Number: CR-3465

Lot Issue Date: 07/28/2017

Expiration Date: 08/31/2020

This Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system. The analyte concentrations were verified by our ISO 17025 accredited laboratory to be within $\pm 5.0\%$, when compared to calibration standards independently prepared using NIST SRM(s). The certified value and uncertainty value at the 95% confidence level for each analyte is determined gravimetrically.


Analyte	True Value			Analytical Method	NIST SRM
boron	10.0	\pm	0.1 $\mu\text{g/mL}$	ICP / ICP-MS	3107
germanium	10.0	\pm	0.1 $\mu\text{g/mL}$	ICP / ICP-MS	3120a
molybdenum	10.0	\pm	0.1 $\mu\text{g/mL}$	ICP / ICP-MS	3134
niobium	10.0	\pm	0.1 $\mu\text{g/mL}$	ICP / ICP-MS	3137
phosphorus	10.0	\pm	0.1 $\mu\text{g/mL}$	ICP / ICP-MS	3139a
rhenium	10.0	\pm	0.1 $\mu\text{g/mL}$	ICP / ICP-MS	3143
silicon	10.0	\pm	0.1 $\mu\text{g/mL}$	ICP / ICP-MS	3150
sulfur	10.0	\pm	0.1 $\mu\text{g/mL}$	ICP / ICP-MS	3154
tantalum	10.0	\pm	0.1 $\mu\text{g/mL}$	ICP / ICP-MS	3155
titanium	10.0	\pm	0.1 $\mu\text{g/mL}$	ICP / ICP-MS	3162a
tungsten	10.0	\pm	0.1 $\mu\text{g/mL}$	ICP / ICP-MS	3163
zirconium	10.0	\pm	0.1 $\mu\text{g/mL}$	ICP / ICP-MS	3169

Matrix: trace nitric and hydrofluoric acids in low TOC water (< 50 ppb)

ULTRA uses purified acids, 18 megohm double deionized water, calibrated Class A glassware & meticulously cleaned bottles in the manufacturing of ULTRAgrade standards. Balances used in the manufacturing of this standard are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001



John Russo
President



Monica Bourgeois
Director of QA/RA